

S E R V I C E N O T E

SUPERSEDES: None

37161A Control Card**Serial Numbers:** 0000A00000 / 9999Z99999**Recommended IC replacement****Duplicate Service Notes:**

37161A-01

37161C-01

37161D-02

To Be Performed By: Agilent-Qualified Personnel or Customer**Situation:**

The IC 1820-3929 is becoming obsolete and will be replaced by 1820-2099. If this part needs to be replaced, use 1820-2099.

DATE: August 1994

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

ENTITY:

ADDITIONAL INFORMATION:

GS

1400