

S E R V I C E N O T E

SUPERSEDES: None

4286A RF LCR Meter

Serial Numbers: 0000A00000 / 9999Z99999

Bit switch setting of the A1 CPU board

To Be Performed By: Agilent-Qualified Personnel

04396-66501/69501 is the CPU board commonly used for 4396A, 4291A and 4286A. 04396-66501/69501 has a bit switch on it. All of its bits are turned off when 04396-66501/69501 is shipped from the factory as a service part. In the case of installing 04396-66501/69501 to 4286A, bit-7 of the bit switch must be turned on. See Figure 1.




Figure-1 CPU board switch setting for 4286A

DATE: January 1996

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
INFORMATION ONLY		
AUTHOR:	ENTITY:	ADDITIONAL INFORMATION:
YM	3355	