

70001A-17A

S E R V I C E N O T E

SUPERSEDES: 70001A-17

HP 70001A Mainframe

Serial Numbers: 0000A00000 / 3330A05901

Exceptions:

3345A05950
3345A05961
3345A05962
3345A05963
3345A05964
3345A05985
3345A05986
3345A05987

Reset Programmable Array Logic (PAL) integrated circuit

Duplicate Service Notes: 70004A-13

Parts Required: 70001-60105

Situation:

With the introduction of the 70500A Real Time Controller the reset PAL on the 70001A A4 (HP-MSIB board) has changed. To accommodate the 70500A the MODRST line on the A4 board will become bi-directional. This will allow modules to assert reset. The new PAL part number is 70001-80032.

Solution/Action:

In the 70001A mainframe replace the A4 board with a 70001-60105. The above serial number exceptions shipped with the 70001-80002 PAL.

DATE: 15 February 1994

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
MODIFICATION AVAILABLE		
ACTION CATEGORY: AGREEABLE TIME		☒ PERFORMANCE ENHANCEMENT ☐ SERVICE/RELIABILITY ENHANCEMENT
LOCATION CATEGORY:	☐ CUSTOMER INSTALLABLE ☐ ON-SITE ☒ HP LOCATION	AVAILABLE UNTIL:
AUTHOR: RDS	ENTITY: 5340	ADDITIONAL INFORMATION: