

S E R V I C E N O T E

SUPERSEDES: 70004A-01A

70004A Display**Serial Numbers:** 0000A00000 / 3545A04550**Firmware Upgrade Kit****To Be Performed By:** Customer or Agilent-Qualified Personnel**Parts Required:** ROM Upgrade Kit, part number 70004-60050**Situation:**

Firmware revisions prior to 7.06, datecode 960424, may experience an intermittent blanking display.

Solution/Action:

Order firmware upgrade kit 70004-60050.

Kit Contents:

- EPROM, Programmed U23
- EPROM, Programmed U24
- EPROM, Programmed U25
- EPROM, Programmed U26
- Installation Note

DATE: January 1997

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:		
MODIFICATION AVAILABLE		
ACTION CATEGORY: AGREEABLE TIME		<input type="checkbox"/> PERFORMANCE ENHANCEMENT <input checked="" type="checkbox"/> SERVICE/RELIABILITY ENHANCEMENT
LOCATION CATEGORY:	<input checked="" type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input type="checkbox"/> SERVICE CENTER	AVAILABLE UNTIL: January 1998
AUTHOR: RDS ENTITY: 5340		ADDITIONAL INFORMATION: