

S E R V I C E N O T E

SUPERSEDES: NONE

70820A Microwave Transition Analyzer

Serial Numbers: 0000A00000 / 3117A00224
Standard Modules Only!

Exceptions: 3177A00180, 3177A00219, 3177A00221, 3177A00222

New Firmware Available, Date Code 911216

To Be Performed By: Agilent-Qualified Personnel or Customer

Situation:

New firmware (rev 1.01) for the 70820A Microwave Transition Analyzer is available. The new firmware will support the 70004A mass storage features including the memory card and GPIB disk access. The GPHIL keyboard interface on the 70004A may also be used for input of active functions.

Many new remote commands have been added including source system commands to support bit error ratio testing. Several limits have also been expanded in this revision of firmware.

Solution/Action:

An upgrade kit P/N 70820-60072 may be ordered. Only standard 70820A's may be upgraded.

Kit Contents:

Installation Note
Rom/Ram Board
Program Guide Programming Supplement

DATE: 21 AUGUST 1992

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR: RDS	ENTITY: 5300	ADDITIONAL INFORMATION:
---------------------------	----------------------------	--------------------------------