

INFORMATION ONLY

8720D-06

S E R V I C E N O T E

Supersedes:
None

8720D Network Analyzer

Serial Numbers: 0000A00000/9999Z99999

Power Levels for Dynamic Range Test

Parts Required:

P/N	Description	Qty.
-----	-------------	------

None

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

Calibration NOT Required

PRODUCT LINE: 5334

AUTHOR: EW

ADDITIONAL INFORMATION: To be performed by Keysight qualified personnel and customers.

Situation:

Power levels for Dynamic Accuracy Test are not documented.

Solution/Action:

Use the following power level values for 8719D/8720D Dynamic Range Test:

8719D/8720D	+5
8719D/8720D Option 007	+10

Revision History:

Service note Revision	Date	Author	Reason For Change
01	Nov 2000	EW	As published