

MODIFICATION RECOMMENDED

**E8363C-07**

**S E R V I C E N O T E**

Supersedes:  
NONE

**E8363C PNA Series Microwave Network Analyzers**

**Serial Numbers: MY49030496, MY49030498**

**Analyzers with the serial numbers listed above must be returned to Agilent for source attenuator(s) replacement.**

**Parts Required:**

<b>P/N</b>	<b>Description</b>	<b>Qty.</b>
33325-60012	source attenuator	1 or 2

**ADMINISTRATIVE INFORMATION**

SERVICE NOTE CLASSIFICATION:		
<b>MODIFICATION RECOMMENDED</b>		
ACTION CATEGORY: AGREEABLE TIME	STANDARDS LABOR: 3.0 Hours	
LOCATION CATEGORY: SERVICE CENTER	SERVICE INVENTORY: SCRAP	USED PARTS: SCRAP
AVAILABILITY: PRODUCT'S SUPPORT LIFE	NO CHARGE AVAILABLE UNTIL: END OF PRODUCT'S SUPPORT LIFE	
AUTHOR: MF	PRODUCT LINE: WN	
ADDITIONAL INFORMATION: Models affected: N5230C, E8362C, E8363C, N5244A, and N5245A. E8363C-07 created 2/13/12.		

© AGILENT TECHNOLOGIES, INC. 2012  
PRINTED IN U.S.A.

February 23, 2012

Rev. 19



**Situation:**

Analyzers with the serial numbers listed on the previous page contain source attenuator(s) that may malfunction.

**Solution/Action:**

- Return the PNA to Agilent for source attenuator(s) replacement.
- Refer to the PNA Service Guide chapter 7 for repair and post-repair procedures. To view the online Service Guide, see the numbered list below for instructions.

Refer to the PNA Service Guide for instructions on removing/replacing microcircuits, post-repair adjustments/procedures, and verification/performance tests. To view this Service Guide information, use the following steps:

1. Go to <http://www.agilent.com>.
2. Enter your PNA model number (Ex: E8363C) in the Search box and click **Search**.
3. Click **Manuals**.
4. Scroll down and click on the title/hyperlink for the Service Guide.
5. When the PDF of the Service Guide is displayed, use the bookmarks in the left column to find the “Removing & Replacing the ... Source Step Attenuators” section in Chapter 7.
6. Follow the instructions for removing and replacing the source attenuator(s).
7. Successful completion of the post-repair adjustments/procedures and verification/performance tests will set the analyzer to its pre-repair condition. All pre-repair certification still applies.

**NOTE:** The PNA contains extremely sensitive components that can be ruined if mishandled. Follow instructions carefully when making cable connections, especially wire harness connections.