A MODIFICATION ONLY - DOES NOT COMMUNICATE

J2126A-09

<u>SERVICE NOTE</u>

Supersedes: NONE

Agilent J2126A Transmission Test Set

Serial Numbers: GB26000000/GB26999999, SG43090100/SG43090136

Replacement of the 2.5Gb/s optics module from a 4-port to a 3-port variant requires the additional selection of a suitable replacement Top Panel.

Parts Required:

P/N Description Qty. J2126-62018 Top Panel, J2126A

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR: BIII McF PRODUCT LINE: PY

ADDITIONAL INFORMATION:

© AGILENT TECHNOLOGIES, INC. 2004 PRINTED IN U.S.A.



February 4, 2004

Page 2 of 2 J2126A-9

Situation:

Older 2.5Gb/s optics modules have four optics ports. Support material has now run out such that these modules are no longer available. The module itself has been replaced by a newer three port variant which has a single multi-rate optical receiver (MRO).

When replacing the 4-port module with a 3-port module, there will be a "spare" hole remaining in the original Top Panel of the instrument.

Solution/Action:

When ordering a new 2.5Gb/s module, you must <u>also</u> order a new **Top Panel** as shown in the table below.

J2126A	Original Material		New Material	
Option	Module	Top Panel	Module	Top Panel
#100	J2126-63001	J2126-65003	J2126-63101	
#101	J2126-63011	J2126-65023	J2126-63111	J2126-62018
#102	J2126-63021	J2126-65033	J2126-63121	





Original Top Panel

New Top Panel