

N5221B-01

Modification Recommended Service Note

Supersedes:
NONE

N5221B – 900 Hz/ 10 MHz to 13.5 GHz PNA Series Microwave Network Analyzer

Serial Numbers: MY60213106, MY60213110, MY60213144

The Problem – A fabrication processing defect in the 26.5 GHz Dual Output microcircuit assembly was recently discovered affecting instruments that were shipped to customers between December 2020 and March 2021. This defect will severely affect the network analyzer’s functionality and should be sent to a local service center at the customer’s earliest convenience.

Parts Required:

P/N	Description	Qty.
N5240-60101	PCA, HMA 26.5 GHz Dual Output	1

ADMINISTRATIVE INFORMATION

ACTION		STANDARDS	
CATEGORY:	<input checked="" type="checkbox"/> AGREEABLE TIME	LABOR: 1.0 Hours	
LOCATION		SERVICE:	USED
CATEGORY:	<input checked="" type="checkbox"/> SERVICE CENTER	INVENTORY:	PARTS: <input checked="" type="checkbox"/> RETURN
AVAILABILITY: PRODUCT'S SUPPORT LIFE		NO CHARGE AVAILABLE UNTIL: September 30, 2022	
<input checked="" type="checkbox"/> Calibration Required		PRODUCT LINE: WN	
		AUTHOR: KS	

ADDITIONAL INFORMATION:

Situation:

A fabrication processing defect in the 26.5 GHz Dual Output microcircuit assembly was recently discovered affecting instruments that were shipped to customers between August 2021 and December 2021. This defect will severely affect the network analyzer’s functionality and should be sent to a local service center at the customer’s earliest convenience.

Solution/Action:

Service Center to replace PCA assembly and calibrate instrument.

Revision History:

Date	Service Note Revision	Author	Reason for Change
13 Sept 2021	01	Keith Stengl	As Published